



of Conformity

Registration No.:

AK 50387916 0001

Report No.:

50080961 001

Holder:

Dirui Industrial Co., Ltd.

95 Yunhe Street

New & High Tech. Development Zone

Changchun, Jilin 130012

P. R. China

Product:

Analysis Equipment

(Urinalysis Hybrid)

Identification:

Type Designation: FUS-3000Plus

(DIRUI)

Serial No.: Engineering Sample

Remark: Refer to test report 50080961 001 for details.

Tested acc. to:

EN 61326-1:2013 EN 61326-2-6:2013 IEC 61326-1:2012 IEC 61326-2-6:2012

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

TÜVRheinlar

Tifizierung'

Date

18.10.2017

Certification Body

Sun Lixun

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg