

Certificate of Analysis

Cat No.	Product Name
744300.4.UC384	STARMag 96 X 4 Universal Cartridge Kit

Lot#	UC9621D13
Date of Manufacture	2021-03-16
Expiry Date	2022-06-15
Package Size	96 rxns X 4 cartridge
Storage	Stored at Room temperature (Universal Proteinase K stored at -20°C)

Components	Specification	Results(PASS/FAIL)
Universal Magnetic beads	DNA recovery / solid content	PASS
Universal Lysis Buffer LB	density / pH	PASS
Universal Binding Buffer BB	density	PASS
Universal Wash Buffer 1 WB1	density	PASS
Universal Wash Buffer 2 WB2	density	PASS
Universal Wash Buffer 3 WB3	density / pH / sterility	PASS
Universal Elution Buffer EB	density / pH / sterility	PASS
Universal Proteinase Buffer PB	density / pH	PASS
Universal Proteinase K	amount / activity	PASS
Appearance	Components Product Name Label Cat No. Lot No. Appearance of box	PASS
Negative Test	Real-Time PCR rxn	PASS
Internal Control Test	Real-Time PCR rxn	PASS

* This Lot was inspected according to Seegene's Quality Control Instruction, and we confirm that all of the results meet the acceptance criteria. This CoA is valid for 1 year from issued date.

(Issued date) 2021-10-13

(Signature)



Head of Quality Assurance Department

Seegene Inc. Taewon Bldg., 91. Ogeum-ro, Songpa-gu, Seoul,
05548, Korea

TEL : (82) 2-2240-4000

FAX : (82) 2-2240-4040

email : cs@seegene.com