

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N° : A SOP N° : TEST-F

Page 1 of 1

CODE	INT2504B	LOT	CE19E09L71	
		Expiry Date	2020-10	

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE19E27L11	Amotosalen	1	180093
assembly	2	2 N/A	2	N/A	
	3	N/A		3	N/A
PCT sub-assembly	1	CE19E09L71	CAD material		90109A / 90111B / 90114C / 90115A

DESCRIPTION: INTERCEPT Platelet Processing Set with Dual Storage Containers

CONFORMITY OF TESTS

	PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)				
Product	Limit (≤)		Date Passed	Results	
Final product	0.125 EU / mL	June 3, 2019		PASS	
Amotosalen solution sub-assembly	☑ 3 mM → 2.0 EU / mL	1	June 3, 2019	PASS	
	OR	2	N/A	N/A	
	□ 6 mM → 0.8 EU / mL	3	N/A	N/A	
PCT sub-assembly	0.25 EU / mL	1	May 13, 2019	PASS	

Disposition : PASS

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09)	Disposition : PASS ☑
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS 🗹
pH ANALYSIS ON CONNECTED KIT (TEST-A)	Disposition : PASS ☑
AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)	Disposition : PASS

STERILITY		
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

4656 UNITS

Reviewed by Fenwal QA MANAGER or DELEGATE Date & Signature

M. Rovina, June 21, 2019

Released by Cerus **QA MANAGER or DELEGATE** Date & Signature

Romain LELIEVRE

June 24th 2019