

## CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT

cerus

Formulaire N°: C22-TEST-F

Révision N° : B

SOP N° : TEST-F

Page 1 of 1

CODE	INT2504B	LOT	CE23C13A01
		Expiry Date	2025-05

		Lot		Ra	w material batch
Amotosalen solution sub-	1	CE23F07L11	Amotosalen	1	220057
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE23C13A01	CAD material		20625B / 20626B / 20721A /
					20722A / 20711B / 20711A /
					20710B / 20723B / 20725B

**DESCRIPTION: INTERCEPT Platelet Processing Set with Dual Storage Containers** 

## CONFORMITY OF TESTS RESULTS

	PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)			
Product	Limits (≤)           0.125 EU / mL         Ju		Date Passed	Results PASS
Final product			e 21, 2023	
Amotosalen solution sub-assembly	☑ 3 mM → 2.0 EU / mL	1	June 26, 2023	PASS
	OR	2	N/A .	N/A
	$\square$ 6 mM $\rightarrow$ 0.8 EU / mL	3	N/A	N/A
PCT sub-assembly	0.25 EU / mL	1	June 21, 2023	PASS

Disposition: PASS

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09)	Disposition : PASS ☑
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS
pH ANALYSIS ON CONNECTED KIT (TEST-A)	Disposition : PASS 🗹
AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)	Disposition : PASS

Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

**QUANTITY** 

**10344 UNITS** 

Reviewed by Fenwal

**QA MANAGER or DELEGATE** Date & Signature

P. MANDRELIER

Released by Cerus QA MANAGER or DELEGATE

Date & Signature

DocuSigned by:

Daniela da Costa Pinto



Signer Name: Daniela da Costa Pinto Signing Reason: I approve this document Signing Time: 7/27/2023 | 2:19:51 AM PDT

E80E5A103DA8483B92F6374B141D762D