

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N° : B

SOP N° : TEST-F

Page 1 of 1

CODE	INT3104B	LOT	CE19G18L71
		Expiry Date	2021-06

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE19H08L11	Amotosalen	1	180093
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE19G18L71	CAD material		81203A1 / 81203A3 / 81210A1

DESCRIPTION: INTERCEPT Processing Set for Plasma

CONFORMITY OF TESTS RESULTS

PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)					
Product	Limits (≤)	Date Passed		Results	
Final product	0.125 EU / mL	August 20, 2019		PASS	
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	August 20, 2019	PASS	
	OR	2	N/A	N/A	
	☑ 6 mM → 0.8 EU / mL	3	N/A	N/A	
PCT sub-assembly	0.25 EU / mL	1	July 24, 2019	PASS	

Disposition: PASS

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09) Disposition: PASS FINAL PHYSICAL TESTING (TEST-C)

Disposition: PASS pH ANALYSIS ON CONNECTED KIT (TEST-A) Disposition: PASS 🗹

AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01) Disposition: PASS

STERILITY		
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑
	Dis	position : PASS

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

2400 UNITS

Reviewed by Fenwal QA MANAGER or DELEGATE

Date & Signature

C. BERTHOMIER

September 12, 2019

Released by Cerus QA MANAGER or DELEGATE Date & Signature

> Romain LELIEVRE Sept 13th 2019

This is to certify that this product was manufactured according to current Good Manufacturing Practice and fulfills the requirements of the Device Master Record.