

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N° : B SOP N° : TE

: TEST-F

Page 1 of 1

CODE	INT3104B	LOT	CE20K24L71
		Expiry Date	2022-10

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE20L14L11	Amotosalen	1	190104
assembly	2	N/A]	2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE20K24L71	CAD material		00720C3 / 01029A1 / 01029B2 /
					01030C3 / 01102C2 / 01102C3

DESCRIPTION: INTERCEPT Processing Set for Plasma

CONFORMITY OF TESTS RESULTS

PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)				
Product	Limits (≤)	Date Passed		Results
Final product	0.125 EU / mL	December 17, 2020		PASS
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	December 22, 2020	PASS
	OR	2	N/A	N/A
		3	N/A	N/A
PCT sub-assembly	0.25 EU / mL	1	November 27, 2020	PASS

Disposition : PASS

✓

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09)	Disposition : PASS 🗹
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS 🗹
pH ANALYSIS ON CONNECTED KIT (TEST-A)	Disposition : PASS
AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)	Disposition : PASS 🗹

STERILITY			
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑	
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑	
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑	
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑	
Disposition · PASS 🗹			

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

2376 UNITS

Reviewed by Fenwal

QA MANAGER or DELEGATE Date & Signature

C. BERTHOMIER

January 6, 2021

Released by Cerus QA MANAGER or DELEGATE Date & Signature

> Romain LÈLIEVRE Jan 07th 2021

This is to certify that this product was manufactured according to current Good Manufacturing Practice and fulfills the requirements of the Device Master Record.