

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N° : B

SOP N° : TEST-F

Page 1 of 1

CODE	INT3104B	LOT	CE20L09L71
		Expiry Date	2022-11

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE20L22L11	Amotosalen	1	200053
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE20L09L71	CAD material		01116C3 / 01116C2

DESCRIPTION: INTERCEPT Processing Set for Plasma

CONFORMITY OF TESTS RESULTS

	PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)				
Product	Limits (≤)	Date Passed		Results	
Final product	0.125 EU / mL	January 5, 2021		PASS	
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	January 5, 2021	PASS	
	OR	2	N/A	N/A	
	☑ 6 mM → 0.8 EU / mL	3	N/A	N/A	
PCT sub-assembly	0.25 EU / mL	1	December 11, 2020	PASS	

Disposition : PASS 🗹

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09)	Disposition : PASS 🗹
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS 🗹
pH ANALYSIS ON CONNECTED KIT (TEST-A)	Disposition : PASS ☑
AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)	Disposition : PASS 🗹

STERILITY		
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑
Disposition : PASS ☑		

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

3240 UNITS

Reviewed by Fenwal

QA MANAGER or DELEGATE

Date & Signature

C. BERTHOMIER

January 20, 2021

Released by Cerus
QA MANAGER or DELEGATE
Date & Signature

Romain LELIEVRE Jan 20th ,2021

This is to certify that this product was manufactured according to current Good Manufacturing Practice and fulfills the requirements of the Device Master Record.