

## CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N°: B SOP N°: T

° : TEST-F

Page 1 of 1

CODE	INT3104B	LOT	CE19F21L72
		Expiry Date	2021-05

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE20G11L11	Amotosalen	1	190048
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE19F21L72	CAD material		81001A1 / 81008A1 / 81015A1

**DESCRIPTION: INTERCEPT Processing Set for Plasma** 

## CONFORMITY OF TESTS RESULTS

	PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)			
Product	Limits (≤)	Date Passed		Results
Final product	0.125 EU / mL	July	/ 17, 2019	PASS
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	July 17, 2019	PASS
	OR	2	N/A	N/A
	☑ 6 mM → 0.8 EU / mL	3	N/A	N/A
PCT sub-assembly	0.25 EU / mL	1	June 25, 2019	PASS

BACTERIOLOGY-BIOBURDEN (TEST-B-177 TEST-B-09)	Disposition: PASS 🗹
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS

pH ANALYSIS ON CONNECTED KIT (TEST-A)

Disposition: PASS

AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)

Disposition: PASS

STERILITY		
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑
	Dier	nocition · PASS 🗹

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

**2208 UNITS** 

Reviewed by Fenwal

QA MANAGER or DELEGATE

Date & Signature

C. BERTHOMIER

April 14, 2020

Released by Cerus
QA MANAGER or DELEGATE
Date & Signature

Romain LELIEVRE April 15th 2020

This is to certify that this product was manufactured according to current Good Manufacturing Practice and fulfills the requirements of the Device Master Record.