

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

Révision N° : B

SOP N° : TEST-F

Page 1 of 1

CODE	INT3104B	LOT	CE19J09L71
		Expiry Date	2021-09

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE19J30L11	Amotosalen	1	190048
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE19J09L71	CAD material		90409C3 / 90415C3

DESCRIPTION: INTERCEPT Processing Set for Plasma

CONFORMITY OF TESTS RESULTS

PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)				
Product	Limits (≤)	Date Passed		Results
Final product	0.125 EU / mL	November 8, 2019		PASS
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	November 6, 2019	PASS
	OR	2	N/A	N/A
		3	N/A	N/A
PCT sub-assembly	0.25 EU / mL	1	October 16, 2019	PASS

Disposition: PASS

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09)	Disposition : PASS 🗹	
FINAL PHYSICAL TESTING (TEST-C)	Disposition : PASS 🗹	

pH ANALYSIS ON CONNECTED KIT (TEST-A)

Disposition: PASS

AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01) Disposition: PASS

STERILITY		
Amotosalen solution	Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
PCT sub-assembly	Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
Connected product	E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
Sterilility Test (TEST-B-13)	No positives	CONFORM ☑
	Dis	position : PASS

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

2568 UNITS

Reviewed by Fenwal

QA MANAGER or DELEGATE

Date & Signature

Released by Cerus
QA MANAGER or DELEGATE

Date & Signature

C. BERTHOMIER

November 20, 2019

Romain LELIEVRE Nov 22nd 2019

This is to certify that this product was manufactured according to current Good Manufacturing Practice and fulfills the requirements of the Device Master Record.