



**JMS SINGAPORE PTE LTD**  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

**Certificate of Analysis**

**PRODUCT CODE : 811-1000**

**LOT NUMBER : 240525002**

**DESCRIPTION : CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE**

| TEST ITEMS           |                        | SPECIFICATION LIMITS | RESULTS |        |
|----------------------|------------------------|----------------------|---------|--------|
| CHEMICAL<br>CONTENTS | pH                     | 5.0 - 6.0            | 5.7     | PASSED |
|                      | CITRIC ACID (%)        | 95 - 105             | 100.8   | PASSED |
|                      | SODIUM BIPHOSPHATE (%) | 95 - 105             | 100.0   | PASSED |
|                      | * TOTAL CITRATE (%)    | 95 - 105             | 100.6   | PASSED |
|                      | DEXTROSE (%)           | 95 - 105             | 99.9    | PASSED |
|                      | ADENINE (%)            | 90 - 110             | 98.6    | PASSED |

**\* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE**

**THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.**

**DATE OF ISSUE : 19-JUL-24**

**CY CHIA**

**QA MANAGER**