



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO **DATE OF ISSUE**
1026696 04-OCT-24

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-4032	BLOOD BAG CPDA-1 (S) 450ML IBSP-NP-CLP 16G CE	240709006	1,040	09/07/24	JUL2027
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	240708001	2,808	08/07/24	JUL2027
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	240711002	1,032	11/07/24	JUL2027

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : BE24/000239

CY CHIA
QA MANAGER



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571178 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jms.com.sg
Reg : 197801579C

Certificate of Analysis

PRODUCT CODE : 811-4032

LOT NUMBER : 240709006

DESCRIPTION : BLOOD BAG CPDA-1 (S) 450ML IBSP-NP-CLP 16G CE

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	98.6	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	100.0	PASSED
	* TOTAL CITRATE (%)	95 - 105	100.3	PASSED
	DEXTROSE (%)	95 - 105	99.2	PASSED
	ADENINE (%)	90 - 110	99.4	PASSED

*** TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE**

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 04-OCT-24

CY CHIA

QA MANAGER