



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO DATE OF ISSUE
1026930 14-DEC-24

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-4012	BLOOD BAG CPDA-1 (S) 450ML CLP 16G CE	241030004	1,000	30/10/24	OCT2027
811-4032	BLOOD BAG CPDA-1 (S) 450ML IBSP-NP-CLP 16G CE	241030006	1,040	30/10/24	OCT2027
814-0132	TRANSFER BAG 150ML CE	241107001	1,000	07/11/24	NOV2027
814-0503	TRANSFER BAG 500ML CE	241104001	2,000	04/11/24	NOV2027
814-0504	TRANSFER BAG 500ML W/Y CONNECTOR CE	241105001	2,080	05/11/24	NOV2027
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	241026001	3,840	26/10/24	OCT2027

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : BE24/000417

CY CHIA
QA MANAGER



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

PRODUCT CODE : 833-8445

LOT NUMBER : 241026001

DESCRIPTION : BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF
16G (DT) CE

CPD

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	99.7	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	100.0	PASSED
	* TOTAL CITRATE (%)	95 - 105	100.7	PASSED
	DEXTROSE (%)	95 - 105	99.0	PASSED
	TOTAL SODIUM (%)	95 - 105	99.2	PASSED

* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

SAGM

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	4.5 - 6.5	5.3	PASSED
	SODIUM (%)	95 - 105	99.6	PASSED
	CHLORIDE (%)	95 - 105	100.0	PASSED
	DEXTROSE (%)	95 - 105	100.6	PASSED
	MANNITOL (%)	95 - 105	101.4	PASSED
	ADENINE (%)	90 - 110	100.2	PASSED

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 14-DEC-24

CY CHIA
QA MANAGER