

FR-QCD-64/B

## **CERTIFICATE OF ANALYSIS**

1	Nar	ne of product	Blood	Bag CPDA-1; 150 ml; Sir	ngle Bag;		
2	Description of product			<b>SINGLE</b> Blood bag with donor tube and donor needle of 16G; with Needle Injury Protector, 1 bag in an Aluminium pack.			
3	Batch No. 19A22Y3				-		
4	Mfg. Date		2022/01				
5	Expiry Date		2024/12				
6	Product Code		PB-1CD156M5S				
7	Batch Size			1860 Nos.			
	Quantitative assay per dosage form						
	Ingredients (g/L)						
8	CP	CPDA-1 Solution – USP		tance Limit	Assay Result		
	а	Monobasic Sodium Phosphate (Monohydrate)	2.11 - 2.33		2.23		
)	b	Dextrose (Monohydrate)	30.30 - 33.50		31.71		
	С	Total Citrate (as Citric acid anhydrous)	19.16 - 21.18		20.71		
	d	Sodium	6.21 - 6.86		6.48		
	е	Adenine	0.247 - 0.303		0.266		
	Other tests or requirements						
	CPDA-1 Solution – USP		Acceptance Limit		Result		
	а	Chloride	Not more than 35 ppm		Passed		
9	b	pH	5.0 - 6.0		5.64		
,	С	Volume of Anticoagulant solution	20 – 23 mL		21		
	d	Bacterial Endotoxin Test	Not more than 5.56 EU/mL		Passed		
	е	Sterility	No growth of microorganisms		Passed		
	f	Particulate matter	Shall comply USP <788>		Passed		
10	The Street	Functional test for PVC container according to ISO 3826-1: 2019		comply with ISO 3826-1	Passed		
Jud	lgme	nt : Complies with USP 43 & ISO 38	326-1:20	)19			
Dat	e: 0	4/02/2022 (DD/MM/YYYY)					
Pre V.V Dep	pare .Kum outy N	ari Alanager (QC) 04/02/02/02/02/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/02/04/04/02/04/04/04/04/04/04/04/04/04/04/04/04/04/	Approved By: Sreejith Nemathillam Analytical Chemist		04/02/2022		

TERUMO PENPOL PRIVATE LIMITED, CIN: U33112KL1985PTC004531, E-Mail: Penpol.info@terumobct.com

Registered Office : I-2, Jawahar Nagar, Kowdiar P.O, Thiruvananthapuram - 695 003 Kerala, India. Phone: +91 471-3015500, +91 471-3015501

Blood Bag Factory : VP 1/520, Puliyarakonam P.O, Thiruvananthapuram - 695 573, Kerala, India. Phone: +91 471-3052022 / 7192210 /3015600 / 3015606 / 3015609

Medical Systems Group: T.C. 27/373, Andoor Buildings, General Hospital Road, Thiruvananthapuram - 695 035 Kerala, India. Phone: +91 471-7135800



FR-QCD-77/A

## **LOT CERTIFICATE**

1	Name of product	Blood Bag CPDA-1; 150 ml; Single Bag;						
2	Description of product	SINGLE Blood bag v	vith donor tube and donor needle of 16G; otector, 1 bag in an Aluminium pack.					
3	Lot No./Batch No.	19A22Y3						
4	Mfg. Date	2022/01						
5	Expiry Date	2024/12						
6	Product Code	PB-1CD156M5S						
7	Lot Size	1860 Nos.	0 Nos.					
8	Intended use	Blood bag is a sterile medical device for single use only and used for collecting whole blood as well as for separation & storage of blood components.						
9	Obtained certificates	CE Mark-Approval of EC Directive 93/42/EEC on medical devices, Annex II (excluding section 4), Certificate no. IN19/818843705 issued by SGS Belgium NV, Notified Body 1639. Quality Management System BSI ISO 9001:2015 certificate number FM747384 issued by BSI Assurance UK Limited, England, UK & IN06/70701 (ISO 13485) issued by SGS United Kingdom Ltd, Under UKAS accreditation.						
10	Type of sterilization	Moist heat sterilization and the sterilization process has been validated and is controlled as specified in ISO-17665-1 with a SAL of 10 <sup>-6</sup> .						
11	Material validation	Quality of plastics is States Pharmacopoeia	validated according to ISO 3826 and United					
	QC specification and inspection result							
	a Chemical test for anticoagulant	solution as per USP	Passed					
12	b Functional test for PVC contain 3826	er according to ISO	Passed					
	c Biological test (Bacterial Endote USP	oxin Test) according to	Passed					
	d Microbiological test (Sterility tes	st) according to USP	Passed					
Judgment : Complies with USP 43 & ISO 3826-1:2019								
Date: 04/02/2022 (DD/MM/YYYY)								
V.V. Dep	Pared By:  Kumari  outy Manager (QC)  OA TO 21  OTHER DESCRIPTION  OUT OF THE PARENT O	Approved Sreejith Ne Analytical	emathillam V V V V V V V V V V V V V V V V V V V					